

ABN: 60 068 057 045

ENZYMATIC TEST KIT FOR THE DETERMINATION OF L-MALIC ACID IN GRAPE JUICE AND WINE FOR DISCRETE ANALYSERS

PRODUCT

Product no.4B160, for in vitro use only.

PRINCIPLE OF MEASUREMENT

L-malic acid is found in grape juice and wine and is determined enzymatically according to the following equations:

MDH ↔ (

Oxaloacetate + NADH + H⁺

L-malic acid is oxidised by nicotinamide adenine dinucleotide (NAD) to oxaloacetate using L-malate dehydrogenase (MDH) enzyme as a catalyst. The equilibrium does not favour formation of oxaloacetate and so oxaloacetate is removed by a trapping enzyme. The amount of NADH formed is measured at 340 nm and is stoichiometrically related to the amount of L-malate consumed. In this method, glutamate oxaloacetate transaminase (GOT) is used as the trapping enzyme. In the presence of L-glutamate, the oxaloacetate is irreversibly converted to L-aspartate.

 $\begin{array}{rcl} GOT \\ Oxaloacetate + L-glutamate & \rightarrow & L-aspartate + \alpha-ketoglutarate \end{array}$

CONTENTS

The kit includes the following reagents:

Reagent No.	Reagent	Quantity	Stability
REAGENT 1	Buffer	19.5mL x 2	24 months at 4°C
REAGENT 2	NAD	5.5mL x 2	12 months at 4°C
REAGENT 3	GOT/MDH	5.5mL x 2	9 months at 4°C

The shelf life of Reagents R1 & R2 can be extended by placing aliquots in a freezer. DO NOT FREEZE Reagent R3 (GOT/MDH).

L-malate + NAD⁺

Failure to store reagents at the recommended temperature will reduce their shelf life.

SAFETY

- Wear safety glasses
- Reagent R1 is mildly corrosive

The reagents contain sodium azide as preservative. Do not ingest.

PROCEDURE

Reagent Definition

Reagent	MA R1	MA R2	MA R3
Stable on board (days)	1	1	1
Alarm limit (mL)	1.0mL	0.7mL	0.7mL
Vial volume	20mL	20mL	20mL
Syringe speed	Normal	Normal	Normal

Test Definition

Test type	Photometric	
Full name	L-Malic acid	
Result unit	g/l	
Number of decimals	2	
Acceptance	Manual	
Dilution 1+	0	

Sample type

Wine, Must, Juice

© Copyright 2018, Vintessential Laboratories. All rights reserved. No part of this publication may be copied or reproduced by any means without the written permission of Vintessential Laboratories.



32 Brasser Avenue Dromana Victoria 3936 Australia T +61 3 5987 2242 F +61 3 5987 3303 E <u>info@vintessential.com.au</u> W <u>www.vintessential.com.au</u>

ABN: 60 068 057 045

Calibration Parameters

Calibration type	Linear	
Repeat time (d)	1	
Points/Calibrator	Duplicate	
Acceptance	Manual	
Curve direction	Ascending	
Type of calibrators	Separate	

Calibrator	Conc. (g/l)	Dil. Ratio 1+
MA 0.00	0.000	0.0
MA 0.10	0.100	0.0
MA 0.50	0.500	0.0
MA 1.00	1.000	0.0
MA 1.50	1.500	0.0
MA 3.00	3.000	0.0

Test Flow



REFERENCES

1. "Compendium of International Methods of Wine and Must Analysis" OIV, Vol 1, 2006, MA-E-AS313-11-ALMENZ, p3.